

Cartilage confined compression test

1. 78 yr old specimens
2. 3 mm thick filter used.
3. Includes PBS, longer rest times between tests (sample kept in fridge during rest) and starting preconditioning and stress relaxation 300 microns off of the 10g find contact position to capture data increase from the instant any contact is established.
4. uniformly thick sample was prepared using vibratome.

day 1: oks00TR5-TBC-LCXX-01-01: thickness: 1.816 mm, positions: 86.818, 86.964.

day 2: oks00TR5-TBC-LCXX-01-02: thickness: 1.818 mm, positions: 48.1385, 48.2315.

day 3: oks00TR5-TBC-LCXX-01-03: thickness: 1.78 mm, positions: 54.1655, 54.2945

